

NANO ZS



Size Distribution Report by Volume

Sample Details

Sample Name: Bodille CS
SOP Name: NarrowHighRes.sop
General Notes:


File Name: CSL-2010-size.dts
Record Number: 762
Material RI: 1.25
Material Absorbtion: 0.01
Dispersant Name: Water
Dispersant RI: 1.330
Viscosity (cP): 0.8872
Measurement Date and Time: Thursday, August 12, 2010 ...

System

Temperature (°C): 25.0
Count Rate (kcps): 130.8
Cell Description: Disposable sizing cuvette
Duration Used (s): 60
Measurement Position (mm): 4.65
Attenuator: 10

Results

	Diam. (nm)	% Volume	Width (nm)
Polydispersity Index: 0.805	Peak 1: 25.71	87.6	4.2
	Peak 2: 171.36	1.8	44.5
	Peak 3: 2951.17	10.6	1071.9



Colloidal Science Laboratory, Inc.

Westampton, NJ 609-267-2065