

NANO ZS



Size Distribution Report by Volume

Sample Details

Sample Name: Oxysilver

SOP Name: CSL-00.sop

General Notes:


File Name:	CSL-2010-size.dts	Dispersant Name:	Water
Record Number:	409	Dispersant RI:	1.330
Material RI:	1.25	Viscosity (cP):	0.8872
Material Absorbtion:	0.01	Measurement Date and Time:	Tuesday, June 29, 2010 12:...

System

Temperature (°C):	25.0	Duration Used (s):	90
Count Rate (kcps):	305.5	Measurement Position (mm):	4.65
Cell Description:	Disposable sizing cuvette	Attenuator:	11

Results

	Diam. (nm)	% Volume	Width (nm)
Polydispersity Index: 0.446	Peak 1: 41.54	33.9	4.4
	Peak 2: 291.22	66.1	30.8
	Peak 3: 0.00	0.0	0.0



Colloidal Science Laboratory, Inc.

Westampton, NJ 609-267-2065