

Analysis Lines

Conditions Sets (All lines share a single condition set)

Sample Introduction

General Settings

Calibration Settings

Standard Concentration

Calibration Parameters

Units


Columns

Display Column Flags


[illegible]

Utopia-ACS-091011. All Data Report 10/11/2009, 1:06:14 PM


Blank (Blk)		10/11/2009, 12:52:29 PM			Tube 1
Label	Sol'n Conc.	Units	SD(Int)	%RSD(Int)	Int. (c/s)
W 207.912	0.000000	ppm	6.894	13.6	50.7288
Ag 328.068	0.000000	ppm	532.224	6.0	8874.39



Silver 30 ppm (Std)		10/11/2009, 12:54:14 PM			Tube 2
Label	Sol'n Conc.	Units	SD(Int)	%RSD(Int)	Int. (c/s)
Ag 328.068	30.0000	ppm	12464.313	1.1	1093568



Tungsten 5 oom (Std)		10/11/2009, 12:56:04 PM			Tube 3
Label	Sol'n Conc.	Units	SD(Int)	%RSD(Int)	Int. (c/s)
W 207.912	5.00000	ppm	169.177	1.3	12802.9



W 207.912 Calibration (ppm)		10/11/2009, 12:56:04 PM			Correlation Coefficient: 1.000000	
Label	Flags	Int. (c/s)	Std Conc.	Calc Conc.	Error	%Error
Blank		50.7288	0.000000	0.000000	-	-
Tungsten 5 oom		12802.9	5.00000	5.00000	0.000000	0.0

Curve Type: Linear

Equation: $y = 2550.4 x + 50.7$

Ag 328.068 Calibration (ppm)		10/11/2009, 12:54:14 PM			Correlation Coefficient: 1.000000	
Label	Flags	Int. (c/s)	Std Conc.	Calc Conc.	Error	%Error
Blank		8874.39	0.000000	0.000000	-	-
Silver 30 ppm		1093568	30.0000	30.0000	0.000000	0.0


Curve Type: Linear

Equation: $y = 36156.5 x + 8874.4$

Utopia-ACS-091011. All Data Report 10/11/2009, 1:06:14 PM

Total Silver (Samp)
10/11/2009, 1:00:07 PM
Tube 4
Weight: 1
Volume: 1
Dilution: 1

Label	Sol'n Conc.	Units	SD	%RSD	Int. (c/s)	Calc Conc.
W 207.912	5.36705	ppm	0.114661	2.1	13739.0	5.36705 ppm
Ag 328.068	26.8763	ppm	0.446602	1.7	980625	26.8763 ppm


Ionic Silver (Samp)
10/11/2009, 1:05:15 PM
Tube 5
Weight: 1
Volume: 1
Dilution: 1

Label	Sol'n Conc.	Units	SD	%RSD	Int. (c/s)	Calc Conc.
W 207.912	3.75243	ppm	0.098646	2.6	9621.05	3.75243 ppm
Ag 328.068	5.15636	ppm	0.044965	0.9	195310	5.15636 ppm

